

Provided by Xinzhu Pu, M.D., Ph.D.
Department of Pharmacology and Toxicology
Indiana University School of Medicine
July 15, 2009

Instrument setting for TEAC assay:

Instrument: Tecan M-200 microplate reader

Plate format: 96-well

Incubation time: 5 min

Reading wavelength: 734 nm